

APR 20 2009

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

|  |   |                      |                         |
|--|---|----------------------|-------------------------|
| Supervisor/Operator Performing the Verification Procedure: |   | DataMaster cdm S/N   | 130121 ✓                |
| Name   | John Pennell                              | ID#                  | 6260 Date 15 April 2009 |
| A Agency   | Fairbanks Airport Police                  | Phone #              | 474-2555                |
| Instrument Location  | 5195 Brumbaugh Ave. Fairbanks, AK 99709 ✓ |                      |                         |
| B Alco S/N   | 57447 ✓                                   | Target Value         | .079 ✓                  |
|  |   | High Pressure        | 975 psi.                |
| Alco Test Values   | 0.080 ✓                                   | 1 <sup>st</sup> Alco | 0.080 ✓                 |
|  |   | 2 <sup>nd</sup> Alco |                         |
| Signature  | John Pennell ✓                            |                      | CRB<br>5/13/09          |
| (OVER)   |   |                      |                         |

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of May, 2009.

Carolyn M. Noland  
(Notary Seal Stamp)

Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

APRIL 15, 2009

OPERATOR'S NAME: PERNELL/JOHN/T  
OPERATOR'S NUMBER: 6260  
SUBJECT'S LAST NAME: VOC  
SUBJECT'S FIRST NAME/MI: VOC  
O.L. #: VOC  
DEPT/AGENCY: FBK6  
CASE/REPORT: VOC  
TEST TYPE: V  
ALCO TARGET VALUE: .079  
ALCO S/N: 57447

### — BREATH ANALYSIS —

.079 ADJUSTED FOR 30.01 in  
ALCO TARGET .079 00:04  
BLANK TEST .000 00:05  
INTERNAL STANDARD VERIFIED 00:05  
ALCO TV 30.01 in .000 ✓ 00:05  
BLANK TEST .000 00:06  
SUBJECT SAMPLE .079 00:07  
BLANK TEST .000 00:08  
ALCO TV 30.01 in .000 ✓ 00:09  
BLANK TEST .000 00:09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

APRIL 15, 2009  
TIME 00:00

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 44c  
BAROMETER: 30.01 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./:0123456789:;<=>?@AB CDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~